


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-14V1-PN-55C1*

S.N.: *7471*

Date: *NOV 10 1995*



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC
250 mA ~~max~~

60 db ATTEN
- 100 VOLTS DIV
PRF ≈ 30 KHz