


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-E4-C-F
S.N.: 7506
Date: DEC 12 1995



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
1 TO 5 NS
- c) Rise time:
 $\leq 0.4 \text{ NS}$
- d) Fall time:
 $\leq 0.4 \text{ NS}$
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz
sl

60 db ATTEN

100 VOLTS/DIV

TOP 1 NS/DIV

BOT 2 NS/DIV

PRF = 100 KHz