


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV14-P.

S.N.: 7546

Date: JAN 16 1996



- a) Output signal amplitude:  
0 TO +30V
- b) Pulse width:  
≤ 1.0 ns
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

40 db ATTN  
10 VOLT/DIV

PRF ≈ 300 kHz

+15 VDC  
200 mA (MAX) MS