


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-P

S.N.: 7547

Date: JAN 16, 1996



a) Output signal amplitude:
0 TO +30 V

b) Pulse width:
 $\leq 1.0 \text{ NS}$

c) Rise time:
NA

d) Fall time:
NA

e) PRF:
0 TO 1.0 MHz

f) Jitter, stability:
OK

g) Prime power:
+ 15 VDC
200 mA (max)

40 dB ATTEN

\therefore 10 VOLTS / DIV

PRF \approx 300 kHz