


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANMP-4-C-N*

S.N.: *7569*

Date: *FEB 27 1996*



- a) Output signal amplitude: *0 TO - 10 VOLTS*
- b) Pulse width: *($R_L = 50\Omega$)
10 NS TO 10 US.*
- c) Rise time: *(10% MAX DUTY CYCLE)
 ≤ 200 PS*
- d) Fall time: *≤ 200 PS*
- e) PRF: *0 TO 100 MHz*
- f) Jitter, stability: *(10% MAX DUTY CYCLE)
OK*
- g) Prime power:

120/240 VOLTS

50 - 60 HZ

[Signature]

TOP *2 NS / DIV
10 VOLTS / DIV (40 dB ATTEN)
PRF = 10 MHz*

Bot *10 US / DIV
10 VOLTS / DIV
PRF = 10 KHz
 $R_L = 50\Omega$*