


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *NH-5A-P*

S.N.: *7574*

Date: *NOV 20 1986*



- a) Output signal amplitude:  
*0 TO +20V*
- b) Pulse width:  
*≤ 200ns*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*40 db ATTEN*  
*10 VOLTS/DIV*

*TOP 500 ps/DIV*

*130V 100 ps/DIV*

*PRF = 300 kHz*

*115 VOLTS*  
*258 mA MAX.*

*[Signature]*