


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV4-5A-T*

S.N.: *7575*

Date: *FEB 20 1996*



- a) Output signal amplitude: *0 to +20V*
- b) Pulse width: *≤ 200 ps*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 to 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VOLTS*
250 mA (MAX)

40 db ATTEN
10 VOLTS/DIV
TOP 500 ps/DIV
BOI 100 ps/DIV
PRF = 300 kHz