


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-5B-PS-PN-EA*

S.N.: *7582*


Date: *APRIL 15 1996*



- a) Output signal amplitude:  
*0 TO ± 500 VOLTS (R<sub>L</sub> ≥ 50Ω)*
- b) Pulse width:  
*0.1 TO 100 μS (1% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 30 NS*

- d) Fall time:  
*≤ 30 NS*
- e) PRF: *0 TO 10 KHz (1% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*

*Ⓐ* Font *R<sub>L</sub> = 50Ω, PRF = 10 KHz*  
*100 VOLTS/DIV*



- g) Prime power:  
*120/240V, 50-60Hz*
- [Signature]*

*Ⓑ* Norm, *R<sub>L</sub> = 50Ω, PRF = 100 Hz*  
*100 VOLTS/DIV.*