


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV14-1A1-PN-554

S.N.: 7597

Date: JAN 30 1996



60 db ATTEN

-100 uV/div

PRF = 30 kHz

a) Output signal amplitude:

+ 100 V

b) Pulse width:

0.6 to 2.0 ns

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 to 100 kHz

f) Jitter, stability:

OK

g) Prime power: +15 VDC

250 mA (max)

h) t_p, t_n : TEMPERATURE

MAINTAINED

[Signature]