


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APP-1-C-P-M-CA-6W*


S.N.: *7609*

Date: *APRIL 26 1986*



- a) Output signal amplitude:
0 TO + 10 VOLTS
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability:
OK
- g) Prime power: *120/240 VOLTS*
50-60 MHz

A
TOP: *1 NS/DIV*
10 VOLTS/DIV
BOT: *100 PS/DIV*
10 VOLTS/DIV



← 10 NS/DIV
10 VOLTS/DIV

← 1 NS/DIV
10 VOLTS/DIV

(40 dB ATTEN)
100 MHz

Dr