


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HVI-PN-SSC1

S.N.: 7614

Date: MAR 21 1986



TOP: P OUT  
BOT: N OUT  
60 dB ATTEN  
100 VOLTS/DIV  
PRF = 30 kHz

- a) Output signal amplitude:  
± 100 VOLTS
- b) Pulse width:  
0.6 TO 2.0 NS
- c) Rise time:  
N/A
- d) Fall time:  
N/A
- e) PRF:  
0 TO 100 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS  
250 mA MAX