


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVPP-1-C-PN

S.N.: 7624

Date: APRIL 22 1996



- a) Output signal amplitude:
0 TO ±10 VOLTS
- b) Pulse width:
0.2 TO 100NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 35 PS
- e) PRF: 0 TO 1.0 MHz


A Out TOP: 1 NS/DIV
130T: 100 PS/DIV
40 dB ATTN, 100 KHz

- f) Jitter, stability:

OK

- g) Prime power:

120/200V
50-60Hz
bb.



B Out TOP: 10 NS/DIV } 40 dB ATTN
130T: 200 PS/DIV } 100 KHz