


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-C-PN*

S.N.: *7631*

Date: *MAR 1 1996*



a) Output signal amplitude:

0 TO ±15V

b) Pulse width:

2 TO 15N

c) Rise time:

≤ 150 PS

d) Fall time:

≤ 600 PS

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

40 dB ATTEN

10 V/V B/PN

PRF ≈ 12 MHz

100 / 200 V/V B
(±10%)

50 - 60 MHz

*USE WITH
SELECTOR CARD
IN 120 V POSITION*