


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMM-2-C-TR-P*

S.N.: *7674*

Date: *JUNE 26 1996*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.5 TO 10 NS
- c) Rise time:
0.3 TO 2.0 NS
- d) Fall time:
0.3 TO 2.0 NS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS

40 dB ATTEN
5 VOLTS/DIV
PRF = 12 MHz
TR 0.3, 0.7 + 2.0 NS

50-60 Hz
[Signature]