


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-U-C-PN-05*

S.N.: *7691*

Date: *JULY 25 1996*



- a) Output signal amplitude: *0 TO ± 50 VOLTS*
- b) Pulse width: *2 TO 100 NS*
- c) Rise time: *≤ 350 PS*
- d) Fall time: *≤ 2 NS*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 Volts*  
*50-60 Hz*  
*AS*

*50 db ATTN*  
*32 VOLTS/DIV*  
*TDR 10 NS/DIV*  
*DOT 1 NS/DIV (RISE TIME)*  
*PRF = 20 KHz*