


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVIR-4-PS P-PLUT -AT*
S.N.: *7696*
Date: *AUG 13 1996*



- a) Output signal amplitude:
0 TO ±200 VOLTS
- b) Pulse width:
RANGE 1: 2 TO 10 NS
RANGE 2: 10 TO 200 NS
- c) Rise time:
RANGE 1: ≤ 1 NS
RANGE 2: ≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power: *120/200 V*
50-60 Hz

60 dB ATTEN
100 VOLTS/DIV
TOP RANGE 1
DOT RANGE 2
PRF ≈ 20 KHz

ALY