


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-4B PW-C-P-0P1


S.N.: 7704

Date: JULY 29 1996



- a) Output signal amplitude:
0 TO +400 V
- b) Pulse width:
0.1 TO 100 μ S
(0.5% MAX DUTY CYCLE)
- c) Rise time:
 ≤ 30 NS
- d) Fall time:
 ≤ 30 NS
- e) PRF: 0 TO 10 KHz
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
170/240 VOLTS
50-60 Hz

Ⓐ $R_L = 50 \Omega$, PRF = 1.0 KHz
100 VOLTS/DIV, 500NS/DIV



Ⓑ $R_L = 50 \Omega$, PRF = 50 KHz
100 VOLTS/DIV, 20 μ S/DIV

[Handwritten signature]