


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-2-P*

S.N.: *7709*

Date: *MAY 13 1996*



- a) Output signal amplitude:  
*0 TO +120 VOLTS*
- b) Pulse width:  
*≤ 2 NS*
- c) Rise time:  
*≤ 1 NS*
- d) Fall time:  
*≤ 1 NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VDC*

*250 mA (max)*

*60 dB ATTN.*

*100 VOLTS/DIV.*

*PRF ≈ 30 KHz*

*[Signature]*