


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-2-T*

S.N.: *7726*

Date: *JUNE 17 1996*



- a) Output signal amplitude:  
*0 TO +120V*
- b) Pulse width:  
*≤ 2NS*
- c) Rise time:  
*≤ 1NS*
- d) Fall time:  
*≤ 1NS*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*60 db ATTEN*  
*100 VOLTS/DIV*  
*PRF = 30 KHz*  
*TOP: 5 NS/DIV*  
*BOT: 2 NS/DIV*

*+15 VDC*  
*300 mA (MAX)*

*[Signature]*