


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-7B-PS-P-UMM*

S.N.: *7741*

Date: *OCT 22 1996*



↑
START

↑
STOP

100 VOLTS/DIV
50 US/DIV
R_L = 100K.

- a) Output signal amplitude:
0 TO +800 VOLTS
- b) Pulse width: *(R_L ≥ 100K)*
100NS TO DC
- c) Rise time:
≤ 50NS
- d) Fall time:
≤ 50NS
- e) PRF:
≤ 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

[Signature]