


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV0-2-C-BE02B-P-R5-M-0P1

S.N.: 7744

Date: JULY 22 1996



- a) Output signal amplitude: 0 TO +2 AMPS
- b) Pulse width: 17 NS
- c) Rise time:  $\leq 2$  NS
- d) Fall time:  $\leq 2$  NS
- e) PRF: 0 TO 20 K/PR
- f) Jitter, stability: OK
- g) Prime power: 170/240 V  
50-60 Hz

*[Signature]*

MONITOR OUT TO 40dB  
AV0-2T-BE02B1 CONNECTED  
VIA 5 METER CABLE.  
1N459A IN SERIES WITH  
1Ω USED AS TEST LOAD  
PRF = 20 KHz