


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-9A-C-P-UTX1*

S.N.: *7746*

Date: *Aug 26 1997*



- a) Output signal amplitude: *0 TO +200 mA*  
*(0 TO +16 V TO 500)*
- b) Pulse width: *0.4 TO 2.0 NS*
- c) Rise time: *≤ 200 ps*
- d) Fall time: *≤ 200 ps*
- e) PRF: *0 TO 100 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*  
*50-60Hz*

*MANUFACTURE OUT*  
*TO 4086*  
*10V/DIV*  
*PRF = 100 kHz*