


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVR14-1-C-PN-07A*

S.N.: *7753*

Date: *Aug 27 1996*



① *250 VOLTS/DIV, 200 NS/DIV*  
*R<sub>L</sub> = 200 Ω*



② *250 VOLTS/DIV, 1.0 NS/DIV*  
*R<sub>L</sub> = 200 Ω*

- a) Output signal amplitude:  
*0 TO ±1000 VOLTS*
- b) Pulse width: *(R<sub>L</sub> = 200 Ω)*  
*0.2 TO 5.0 μS*
- c) Rise time:  
*≤ 50 NS*
- d) Fall time:  
*≤ 50 NS*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power: *120/240 V*  
*50-60 Hz.*

*[Signature]*