


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NR-4A-PW-C-P*

S.N.: *7759*

Date: *SEPT 27 1996*



- a) Output signal amplitude: *0 TO +400 VOLTS*
- b) Pulse width: *($R_L \geq 50\Omega$)*
50 NS TO 5.0 US
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
20 NS

d) Fall time:
20 NS


e) PRF: *0 TO 10 KHz*
(0.5% MAX DUTY CYCLE)

f) Jitter, stability:
OK

g) Prime power: *100 ± 10% V*
50-60 Hz.

(A)

$R_L = 50\Omega$ PRF = 5 KHz
100 V/DIV



$R_L = 50\Omega$, PRF = 1 KHz
100 V/DIV
(WIDE PULSE)

[Signature]