


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANIMK-1-C-P-AN-F14A*

S.N.: *7770*

Date: *OCT 17 1996*



- a) Output signal amplitude: *0 TO ± 5 VOLTS*
- b) Pulse width: *20 TO 200 NS*
- c) Rise time: *≤ 1 NS*
- d) Fall time: *≤ 1 NS*
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

(A)

40 dB ATTEN
.5 VOLTS/DIV

TOP 5 NS/DIV

BOOT 20 NS/DIV

PRF = 50 KHz

[Signature]