


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MG-4B-N-P5-151*

S.N.: *7799*

Date: *DEC 20 1996*



- a) Output signal amplitude:
-200 TO -850 VOLTS
- b) Pulse width: ($R_L = 50 \Omega$)
6 NS
- c) Rise time:
 $\leq 2 NS$
- d) Fall time:
 $\leq 3 NS$
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 Hz

60 db ATTN
100 VOLTS/DIV
5 NS/DIV
PRF = 1 KHz

[Signature]