


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-E2-15-P-0A-EW-05*

S.N.: *7807*

Date: *NOV 13 1996*



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
70 TO 200 NS
- c) Rise time:
60.5 NS
- d) Fall time:
60.5 NS
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/200V
52-60Hz

50 db ATTN

- = 32 V/3/DIV

TOP 20 NS/DIV

BOT 1 NS/DIV
(RISE TIME)

PRF = 50 kHz