


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-155B-PS-N-NAPP1*

S.N.: *7809*

Date: *OCT 9 1996*



- a) Output signal amplitude:
PULSE: 0 TO -200 mA
- b) Pulse width:
OFFSET: 0 TO -200 mA
100 NS TO 10 US
- c) Rise time:
≤ 50 NS
- d) Fall time:
≤ 50 NS
- e) PRF:
0 TO 1 MHz.
- f) Jitter, stability:

OK

- g) Prime power:
1.20 / 240 V
50 - 60 Hz.

RESISTIVE LOAD, $R_L = 1 \Omega$
500 NS / DIV.

TOP MONITOR OUT TO
MULTI 2, 200 mV / DIV

BOT LOAD VOLTAGE.
500 mV / DIV

PRF = 10 KHz.