


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AN-144B1*

S.N.: *7829*

Date: *DEC 9 1996*



- a) Output signal amplitude: *+20 VOLTS TO*
- b) Pulse width: *50 n*  
*10 NS TO DC*
- c) Rise time: *≤ 2 NS*
- d) Fall time: *≤ 2 NS*
- e) INPUT: *TTL.*
- f) Jitter, stability: *OK*
- g) Prime power: *+28 V DC*  
*450 mA (MAX)*

*40 db ATTN*

*10 VOLTS/DIV.*