


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-P.  
S.N.: 7835  
Date: DEC 5 1996



- a) Output signal amplitude:  
0 TO +30V
- b) Pulse width:  
 $\leq 1NS$
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VDC  
150 mA (MAX)

*[Signature]*

40dB ATTEN  
10 V/DIV  
PRF = 100 kHz