


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PA-FD1-PS*

S.N.: *7851*

Date: *JAN 13 1997*



- a) Output signal amplitude:
TTL, $R_L \geq 50\Omega$
- b) Pulse width:
50ns to 50 μ s
- c) Division factor:
2 to 999
- d)
- e) PRF: *0 to 10 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz

TOP INPUT, ≈ 10 MHz

BOT OUT, $\div 10$

$R_L = 50\Omega$

[Signature]