


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMN-1-C-P

S.N.: 7863

Date: JAN 27 1997



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
0.3 TO 5.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 50 MHz
- f) Jitter, stability:
OK

g) Prime power:
100/200 V
± 10 %

[Handwritten signature]

40 db ATTN
5 VOLTS/DIV

TOP 500 PS/DIV
PRF ≈ 12 MHz

MIDDLE 1.0 NS/DIV
PRF ≈ 50 MHz

BOTTOM AS MIDDLE
5.0 NS/DIV