


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVS-2-C-BE03-R5P-0P1-M*

S.N.: *7886*

Date: *JAN 20 1997*



- a) Output signal amplitude:  
*0 TO +2 AMP*
- b) Pulse width:  
*17 NS*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*MONITOR OUT TO 4066  
(100V/DIV, 2 AMP/DIV,  
5 NS/DIV).*

*AVS-2T-BE02B3 CONNECTED  
VIA 5 METAL CABLE  
IN 459A IN SERIES WITH  
1Ω USED AS TEST  
LOAD. PRF = 20 KHz.*