


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-2-C.-BE03-R5-P-0P1-M*

S.N.: *7887*

Date: *JAN 31 1997*



- a) Output signal amplitude:
0 TO +2 AMP
- b) Pulse width:
17 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz.
- f) Jitter, stability:
OK

g) Prime power: *120/240V*
50-60 Hz

MONITOR OUT TO 40db
100V/DIV, 2AMP/DIV
5NS/DIV

AVO-2T-BE02B3 CONNECTED
VIA A 5 METER CABLE
1A459A IN SERIES WITH
1-2 4500 OH TEST
LOAD. PRF = 20 KHz

[Signature]