


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVNN-1-e-STA*

S.N.: *7955*

Date: *APRIL 9 1997*



a) Output signal amplitude:
0 TO 10 VOLTS (PP)

b) Pulse width:

c) Rise time:
≤ 150 ps

d) Fall time:
≤ 150 ps

e) PRF: *25 TO 100 MHz*

f) Jitter, stability:
OK

g) Prime power:
120/240V
540-60dB
[Signature]

40 db ATTEN
10 VOLTS / DIV

TRIG 5 NS / DIV (≈ 100 MHz)

MX10 500 ps / DIV (RISE + FALL)

DOT 10 NS / DIV (≈ 25 MHz)