


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMM-1-P*

S.N.: *7956*

Date: *JAN 30 1997*



- a) Output signal amplitude:
0 TO 70 Volts
- b) Pulse width:
200 ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+24VDC

400mA max

40dB ATTEN

-10 VOLT/DIV

PRF ≈ 10 MHz

[Signature]