


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMMH-1-P*

S.N.: *7958*

Date: *JAN 30 1997*



a) Output signal amplitude:
0 TO +10 VOLTS

b) Pulse width:
≤ 200 ns.

c) Rise time:
NA

d) Fall time:
NA

e) PRF: *0 TO 25 MHz.*

f) Jitter, stability:
OK

g) Prime power:
+24 VDC.

400 mA MAX
[Signature]

40 dB ATTEN
10 VOLTS/DIV
PRF ≈ 10 MHz.