


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-1-P*

S.N.: *7960*

Date: *FEB 4 1997*



- a) Output signal amplitude:
0 to +10 V
- b) Pulse width:
≤ 200 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 dB ATN

10 V/DIV

PRF = 10 MHz

+ 24 VDC
40 mA (MAX)
MF