


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVMMH-1-P*

S.N.: *7963*

Date: *FEB 5 1997*



- a) Output signal amplitude:  
*0 TO +10V*
- b) Pulse width:  
*≤ 200pS*
- c) Rise time:  
*✓ NA*
- d) Fall time:  
*✓ NA*
- e) PRF:  
*✓ 0 TO 25 MHz*
- f) Jitter, stability:  
*✓ OK*
- g) Prime power:

*40 dB ATTEN*  
*10 VOLTS / DIV*  
*PRF = 25 MHz*

*+24 VDC*

*400 mA (max)*