


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANM4-1-P*

S.N.: *7965*

Date: *FEB 5 1997*



a) Output signal amplitude:

*0 TO +10V*

b) Pulse width:

*≤ 200ns*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+24 VDC*

*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF ≈ 25 MHz*

*400mV (max)*