


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVP-25-P-PS-EA

S.N.: 7979

Date: APRIL 10 1997



- a) Output signal amplitude:  
0 TO +2 VOLTS
- b) Pulse width:  
0.2 TO 4.0 NS
- c) Rise time:  
≤ 40 PS
- d) Fall time:  
≤ 135 PS
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240V  
50-60 Hz.

20 dB ATTEN  
∴ 2 VOLTS / DIV  
PRF = 1.0 MHz

$$T_{RT} = \sqrt{T_{R25}^2 + T_{SCOPE}^2 + T_{ATTEN}^2}$$

≈ 50 PS (20-80%)

$$T_{SCOPE} = 25 PS$$

18 GHz ATTEN + CABLES

∴ CONCL THAT  $T_{R25} \leq 40 PS$

*[Handwritten signature]*