


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV-155C-C-N*

S.N.: *7982*


Date: *APRIL 21 1999*



- a) Output signal amplitude: *PULSE: 0 TO -2.0 AMPS*
DC: 0 TO -500mA
- b) Pulse width: *100 NS TO 10US.*
- c) Rise time: *≤ 50NS*
- d) Fall time: *≤ 50NS*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*
50-60Hz.

TOP MONITOR OUT TO 10K
1 VOLT/DIV

BOT LOAD RESISTOR VOLTAGE
1V/DIV



[Handwritten signature]