


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-U-0-P*

S.N.: *7986*

Date: *MAY 8 1997*



- a) Output signal amplitude: *0 TO +50 VOLTS*
- b) Pulse width: *2 TO 100 NS*
- c) Rise time: *≤ 35 ps.*
- d) Fall time: *≤ 2 ns*
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz.*

*50 db ATTEN*  
*32 VOLTS/DIV*  
*TOP: 10 NS/DIV*  
*BOT: 0.5 NS/DIV*  
*(RISE TIME)*