


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-AV-15 - P5 P*

S.N.: *8002*

Date: *AUG 23 1994*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0N*
- c) Rise time:  
*≤ 60 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*

*40 dB ATTEN*  
*5 VOLTS/DIV*  
*PRF = 1.0 MHz*

*50-60 Hz*  
*[Signature]*