


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-E3-W-P5-P*

S.N.: *8007*

Date: *JUNE 24 1997*



- a) Output signal amplitude:  
*0 TO +100V*
- b) Pulse width: ( $R_L = 50\Omega$ )  
*50 NS TO 5.045*
- c) Rise time:  
 *$\leq 0.5 NS$*
- d) Fall time:  
 *$\leq 1.0 NS$*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 Vrms*  
*50-60 Hz*

*60 dB ATTN*  
*100 V/DIV*  
*TOP 10 NS/DIV*  
*BUT 1 NS/DIV*  
*(RISE TIME)*