




PULSE GENERATOR
PERFORMANCE CHECK

Model: *MN-151-C*

S.N.: *8016*

Date: *SEPT 26 1984*



- a) Output signal amplitude: *0 TO ± 200 VOLTS*
- b) WAVEFORMS: *SINE
SQUARE
TRIANGLE
DC OFFSET*
- c) Rise time: } *1 µS RT*
- d) Fall time: } *MIN Volt.*
- e) PRF: *10 Hz TO
100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *170/2000 V
50-60Hz*