


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-3AH-C-P-M

S.N.: 3041

Date: OCT 4 1994



a) Output signal amplitude:  
0 TO +450 VOLTS

b) Pulse width: (at 50%)  
5 TO 100 NS

c) Rise time:  
 $\leq 1$  NS

d) Fall time:  
 $\leq 4$  NS

e) PRF:  
2 NS/DIV 0 TO 5 KHz

f) Jitter, stability:  
OK

g) Prime power:  
12 / 240 VOLTS  
50-60 Hz

70 db ATTEN  
320 VOLTS/DIV  
PAF = 2 KHz

*[Signature]*