


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-C-P*

S.N.: *8044*

Date: *OCT 19 1994*



- a) Output signal amplitude:
0 TO +5 VOLTS
- b) Pulse width:
10 TO 200 NS
- c) Rise time:
(20% MAX DUTY CYCLE)
≤ 1 NS
- d) Fall time:
≤ 1 NS
- e) PRF: *0 TO 10 MHz*
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 VOLTS
50 - 60 MHz

TOP: 50 NS/DIV, 5 VOLT/DIV
BOT: 5 NS/DIV, 5 VOLT/DIV

P_{cc} = 50 W

[Signature]