


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MVH-HV1-P*

S.N.: *8092*

Date: *JUNE 2 1997*



- a) Output signal amplitude: *0 TO +100V*
- b) Pulse width: *≤ 10NS*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 VDC*

I_S ≤ 250 mA.

[Signature]

*60 db ATTEN
100 VOLTS DIV
PULSE ≈ 30 KHz.*