


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AN6-1-F*
S.N.: *3093*
Date: *JUNE 2 1997*



- a) Output signal amplitude:
0 TO +250 V
- b) Pulse width:
≤ 4 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN
100 V OVER 10V
PRF ≈ 30 KHz

+15 VDC
350 mA (MAX)
[Signature]